

K052884

OCT 26 2005

510(k) SUMMARY

1. **DEVICE NAME:** Solid State X-Ray Imager
Model Name: DFP-8000D / FPD
Trade/Proprietary Name: Digital Radiography System with Flat Panel Detector (FPD)
2. **ESTABLISHMENT REGISTRATION:** 2020563
3. **U.S. Agent Name and Address:** TOSHIBA AMERICA MEDICAL SYSTEMS, INC.
2441 MICHELLE DRIVE
TUSTIN, CA 92780
Contact Person: Michaela Mahl
Senior Regulatory Affairs Specialist
(714) 730 - 5000
4. **Manufacturing Site:** TOSHIBA MEDICAL SYSTEMS CORPORATION
1385 Shimoishigami
Otwara-shi, Tochigi 324-8550, Japan
5. **Date of Submission:** October 05, 2005
6. **Predicate Device:** DFP-8000D with Image Intensifier (K013608)
7. **DEVICE DESCRIPTION**
This equipment is a digital radiography with Flat Panel Detector (FPD) used in diagnostic X-ray angiography system configuration.
This equipment processes, displays, and records digital images obtained from the Flat Panel Detector, and it replays the recorded images.
8. **SUMMARY of INTENDED USE**
This device is a digital radiography used in diagnostic X-ray angiography system configuration.

This X-ray angiography system is indicated for use in diagnostic and interventional procedures for cardiac blood vessels, cerebral blood vessels, abdominal blood vessels and lower limb blood vessels.

It is intended to replace images obtained through the image intensifier technology.

9. EQUIVALENCY INFORMATION

TOSHIBA Medical Systems Corporation believes that the new Digital Radiography System, model DFP-8000D/FDP is substantially equivalent to the current Digital Radiography System, model DFP-8000D with Image Intensifier (model RTP16301J-G1E) (K013608) except for the new Flat Panel Detector (FPD).



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

OCT 26 2005

Toshiba America Medical Systems, Inc,
c/o Mr. Mark Job
Responsible Third Party Official
Regulatory Technology Services LLC
1394 25th Street, NW
BUFFALO MN 55313

Re: K052884
Trade/Device Name: Digital Radiography System with Flat Panel Detector
Model DFP-8000D/FPD
Regulation Number: 21 CFR §892.1650
Regulation Name: Image-intensified fluoroscopic x-ray system
Regulatory Class: II
Product Code: MQB
Dated: October 10, 2005
Received: October 13, 2005

Dear Mr. Job:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at one of the following numbers, based on the regulation number at the top of this letter:

21 CFR 876.xxxx	(Gastroenterology/Renal/Urology)	240-276-0115
21 CFR 884.xxxx	(Obstetrics/Gynecology)	240-276-0115
21 CFR 892.xxxx	(Radiology)	240-276-0120
Other		240-276-0100

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Nancy C. Brogdon
Director, Division of Reproductive,
Abdominal, and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K052884

Device Name: Digital Radiography System with Flat Panel Detector, Model DFP-8000D/FDP

Indications for Use:

This device is a digital radiography system used in diagnostic X-ray angiography system configuration.

This X-ray angiography system is indicated for use in diagnostic and interventional procedures for cardiac blood vessels, cerebral blood vessels, abdominal blood vessels and lower limb blood vessels.

It is intended to replace images obtained through the image intensifier technology.

Prescription Use ☒
 (Part 21 CFR 801 Subpart D)

OR

Over-The-Counter Use ☐
 (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Nancy C Broglon
(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
 510(k) Number K052884